


Declaration of Conformity

In accordance to EN ISO/IEC 17050-1:2010

We, Esco Micro Pte. Ltd.
Of 21 Changi South Street 1
Singapore, 486777
Tel: +65 6542 0833



declare on our sole responsibility that the product

Category : Laboratory Fume Hood
Brand : Frontier® Perchloric Acid
Model : EFP-4UD4VW-, EFP-4UD4CW-, EFP-4UD6VW-, EFP-4UD6CW-, EFP-5UD4VW-,
EFP-5UD4CW-, EFP-5UD6VW-, EFP-5UD6CW-, EFP-6UD4VW-, EFP-6UD4CW-,
EFP-6UD6VW-, EFP-6UD6CW-, EFP-8UD4VW-, EFP-8UD6VW-

in accordance with the following directives:

2014/35/EU : The Low Voltage Directive and its amending directives
2014/30/EU : The Electromagnetic Compatibility Directive and its amending directives

has been designed to comply with the requirement of the following Harmonized Standard:

Low Voltage : EN 61010-1:2010
EMC : EN 61326-1:2013 Class B
Design/Performance Criteria : ANSI/ASHRAE 110-2016 and European Union Standard EN 14175-3

More information may be obtained from Esco Lifesciences' authorized distributors located within the European Union. A list of these parties and their contact information is available upon request.

Place of Issue: Singapore

Date of Issue: December 2022

Indra Setiawan

Indra Setiawan

Plant Manager, Esco Lifesciences

GLOBAL HEADQUARTERS

Singapore, Indonesia, Philippines

GLOBAL OFFICES

ASEAN: Malaysia, Thailand, Vietnam | **FAR EAST:** China, Hong Kong, Taiwan, South Korea, Japan | **SOUTHASIA:** Bangladesh, India | **MIDDLE EAST:** UAE
AFRICA: South Africa | **NORTH AMERICA:** USA | **EUROPE:** United Kingdom, Italy, Denmark, Lithuania, Russia