

# CERTIFICATE

It is hereby confirmed that the

## **Fume Cupboard Bench Type Frontier Acela 6ft.**

of the manufacturer

**Esco Micro Pte Ltd  
21 Changi South Street 1  
Singapore 486777**

was tested according to

# **EN 14175 part 3**

at the total volume flow rate of **871 m<sup>3</sup>/h**

The details and data of the complete test procedure  
are content of the type test report

**ESCO/EN/03/2008**

which is given to the manufacturer in

25<sup>th</sup> August 2008

Certified by:  
Tintschl Engineering AG  
Goerdelerstr. 21  
91058 Erlangen  
Germany

.....  
Dipl.-Ing. Bernd Schubert  
Erlangen, / /

