

## Declaration of Conformity

In accordance to EN ISO/IEC 17050-1:2010



We, Esco Micro Pte. Ltd.  
Of 21 Changi South Street 1  
Singapore, 486777  
Tel: +65 6542 0833

declare on our sole responsibility that the product

**Category** : Laboratory Fume Hood  
**Brand** : Frontier® Perchloric Acid  
**Model** : EFP-4UD4VW-, EFP-4UD4CW-, EFP-4UD6VW-, EFP-4UD6CW-, EFP-5UD4VW-,  
EFP-5UD4CW-, EFP-5UD6VW-, EFP-5UD6CW-, EFP-6UD4VW-, EFP-6UD4CW-,  
EFP-6UD6VW-, EFP-6UD6CW-, EFP-8UD4VW-, EFP-8UD6VW-

in accordance with the following directives:

**Electrical Equipment (Safety) Regulations 2016**  
**Electromagnetic Compatibility Regulations 2016**

has been designed to comply with the requirement of the following Harmonized Standard:

**Low Voltage** : EN 61010-1:2010  
**EMC** : EN 61326-1:2013 Class B  
**Design/Performance Criteria** : ANSI/ASHRAE 110-2016 and European Union Standard EN 14175-3

More information may be obtained from Esco Lifesciences' authorized distributors/local offices. A list of these parties and their contact information is available upon request.

Place of Issue: Singapore

Date of Issue: December 2022

*Indra Setiawan*

**Indra Setiawan**

Plant Manager, Esco Lifesciences

---

**GLOBAL HEADQUARTERS**

Singapore, Indonesia, Philippines

**GLOBAL OFFICES**

**ASEAN:** Malaysia, Thailand, Vietnam | **FAR EAST:** China, Hong Kong, Taiwan, South Korea, Japan | **SOUTHASIA:** Bangladesh, India | **MIDDLE EAST:** UAE  
**AFRICA:** South Africa | **NORTH AMERICA:** USA | **EUROPE:** United Kingdom, Italy, Denmark, Lithuania, Russia