

Declaration of Conformity

In accordance to EN ISO/IEC 17050-1:2010



We, Esco Micro Pte. Ltd.
Of 21 Changi South Street 1
Singapore, 486777
Tel: +65 6542 0833

declare on our sole responsibility that the product

Category : Laboratory Fume Hood
Brand : Frontier® Acid Digestion™
Model : EFQ-4UDCCW-_, EFQ-4UDPCW-_, EFQ-4UDCVW-_, EFQ-4UDPVW-_, EFQ-5UDCCW-_,
EFQ-5UDPCW-_, EFQ-5UDCVW-_, EFQ-5UDPVW-_, EFQ-6UDCCW-_, EFQ-6UDPCW-_,
EFQ-6UDCVW-_, EFQ-6UDPVW-_, EFQ-8UDCVW-_, EFQ-8UDPVW-_

in accordance with the following directives:

Electrical Equipment (Safety) Regulations 2016
Electromagnetic Compatibility Regulations 2016

has been designed to comply with the requirement of the following Harmonized Standard:

Low Voltage : EN 61010-1:2010
EMC : EN 61326-1:2013 Class B
Design/Performance Criteria : ANSI/ASHRAE 110-2016 and European Union Standard EN 14175-3

More information may be obtained from Esco Lifesciences' authorized distributors/local offices. A list of these parties and their contact information is available upon request.

Place of Issue: Singapore

Date of Issue: December 2022

Indra Setiawan

Indra Setiawan

Plant Manager, Esco Lifesciences

GLOBAL HEADQUARTERS

Singapore, Indonesia, Philippines

GLOBAL OFFICES

ASEAN: Malaysia, Thailand, Vietnam | **FAR EAST:** China, Hong Kong, Taiwan, South Korea, Japan | **SOUTHASIA:** Bangladesh, India | **MIDDLE EAST:** UAE
AFRICA: South Africa | **NORTH AMERICA:** USA | **EUROPE:** United Kingdom, Italy, Denmark, Lithuania, Russia