

CERTIFICATE OF ANALYSIS



This Certificate is issued to Applicant:

Esco Micro Pte Ltd
19 Changi South Street 1
Singapore 486779

Add value.
Inspire trust.

FOR

Name: **Labculture® BSC, Class II, LA2-G G4 GMP Model Biosafety Cabinet**

Product:	LA2-3G8 G4	LA2-3G8-S4 G4	LA2-3G8-S6 G4
Models:	LA2-4G8 G4	LA2-4G8-S4 G4	LA2-4G8-S6 G4
	LA2-5G8 G4	LA2-5G8-S4 G4	LA2-5G8-S6 G4
	LA2-6G8 G4	LA2-6G8-S4 G4	LA2-6G8-S6 G4

Manufacturer: Esco Micro Pte Ltd

Specification: ISO Class 4 in Table 1, ISO 14644-1: 2015

Test Reports: 7191358783-CHM25-YL, 7191358783-CHM25-YL-R1

Tested according to: ISO 14644-1: 2015, ISO 14644-14: 2016

SUMMARY

The product (Name: **Labculture® BSC, Class II, LA2-G G4 GMP Model Biosafety Cabinet; Model: LA2-6G8-G4**) had been tested and was found to comply with the specification cleanroom compatibility stated above. This Summary refers to the sample as stated in above product table. Quality or safety of the sample is not included in this Summary.

DR YANG LEI
EXECUTIVE CONSULTANT
ELEMENTAL / ENVIRONMENTAL ANALYSIS
CHEMICAL & MATERIALS

Reference report No.: 7191358783	Date of Original Issue: 6 JUN 2025	Date of Last Review: 22 JUL 2025	Suggest Next Review Date: 5 JUN 2027
----------------------------------	------------------------------------	----------------------------------	--------------------------------------

This Summary is part of a full report and should be read in conjunction with it. This Summary remains the property of TÜV SÜD PSB Pte Ltd and shall be returned upon request. The use of this Summary is subjected to the terms and conditions of the Test Report.



TÜV SÜD PSB

Laboratory:
TÜV SÜD PSB Pte. Ltd.
15 International Business Park
TÜV SÜD @ IBP
Singapore 609937

Phone : +65-6778 7777
E-mail: info.sg@tuvsud.com
<https://www.tuvsud.com.sg>
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
15 International Business Park
TÜV SÜD @ IBP
Singapore 609937
TÜV