

EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms.

In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

<i>Product designation:</i>	Laboratory centrifuge		
<i>Product name:</i>	Sigma 1-7		
<i>Order number:</i>	100011		
<i>Directives:</i>	2006/42/EC	Machinery Directive	
	2014/35/EU	Low Voltage Directive	
	2014/30/EU	EMC Directive	
	(EU) 2015/863	RoHS Directive	
<i>Normes:</i>	EN 61010-2-020:2017 EN IEC 61326-1:2021		

Sigma Laborzentrifugen GmbH

An der Unteren Söse 50
37520 Osterode
Germany

Authorised representative
for CE matters:
Eckhard Tödteberg

Osterode, 2024-12-04

A handwritten signature in black ink that reads "Michael Sander".

Managing Director